

Requirements:

- Urea/Ammonia (Rapid) Assay Kit (K-URAMR) (provides ~ 600 assays).
- K-URAMR (UREA) ChemWell®-T assay file.
- Use in association with the Urea/Ammonia (Rapid) Assay Kit (K-URAMR) product data booklet.



Use:

For the specific measurement of urea especially in wines, fruit juices, beverages and food products.

For specific sample preparation methods refer to the Urea/Ammonia (Rapid) Assay Kit (K-URAMR) data booklet.

Assay Principle:

Conversion of urea via the following reactions is directly proportional to the coupled consumption of NADPH:



Procedure:

Prepare the assay reagents and calibrators and use with the K-URAMR (UREA) ChemWell®-T assay file.

Urea/Ammonia (Rapid) Assay Kit Components:

Bottle 1: Buffer (18 mL, pH 8.0) plus 2-oxoglutarate and sodium azide (0.02% w/v) as a preservative. Stable for > 2 years at 4°C.

Bottle 2: NADPH. Stable for > 5 years at -20°C.

Bottle 3: Glutamate dehydrogenase suspension (1.1 mL). Stable for > 2 years at 4°C.

Bottle 4: Urease solution (2.7 mL). Stable for > 2 years at -20°C.

Bottle 5: Urea control powder (~ 2 g). Stable for > 2 years at room temperature.

Preparation of Kit Components:

- 1 & 3. Use the contents of bottles 1 and 3 as supplied. Stable for > 2 years at 4°C.
2. Dissolve the contents of bottle 2 in 12 mL of distilled water. **Stable for > 1 year at 4°C** or > 2 years at -20°C (to avoid repetitive freeze / thaw cycles, divide into appropriately sized aliquots and store in polypropylene tubes).
4. Use the contents of bottle 4 as supplied. Stable for > 2 years at -20°C.
5. To prepare the K-URAMR Urea 4 calibrator (0.1 g/L) dissolve 100 mg of urea in 1 L of distilled water. This solution is stable for ~3 months at -20°C.

Use serial dilutions of K-URAMR Urea 4 to prepare K-URAMR Urea 2 and K-URAMR Urea 3.

Preparation of Assay Reagents: (per ~ 100 assays)

Reagent 1:

Component	Volume
distilled water	15 mL
bottle 1 (buffer)	3 mL
*bottle 2 (NADPH)	2 mL
Total volume	20 mL

*after adding 12 mL of distilled water

Reagent 1 stability: > 7 days at 4°C

Reagent 2:

Component	Volume
distilled water	2.4 mL
bottle 3 (GIDH)	0.18 mL
Total volume	2.58 mL

Reagent 2 stability: > 7 days at 4°C

Reagent 3:

Component	Volume
distilled water	2.1 mL
bottle 4 (urease)	0.45 mL
Total volume	2.55 mL

Reagent 3 stability: > 2 days at 4°C

Calibrators:

- K-URAMR Urea 1: 0 g/L (use distilled water)
- K-URAMR Urea 2: 0.025 g/L urea
- K-URAMR Urea 3: 0.05 g/L urea
- K-URAMR Urea 4: 0.1 g/L urea

Assay Parameters:

- Assay volumes: Reagent 1: 0.200 mL
Sample: 0.010 mL
Reagent 2: 0.025 mL
Reagent 3: 0.025 mL
- Calibrators: 0, 0.025, 0.05, 0.1 g/L urea
- Reaction time: 5 min at 37°C plus 6 min at 37°C
- Wavelength: 340 nm
- Assay type: endpoint
- Reaction direction: decrease
- Linearity: up to 0.1 g/L of urea

