

Megazyme

MICROPLATE FORMAT
ASSAY PROCEDURE
For

ACETIC ACID
(Rapid, Manual, Simple and
End-Point AK / PTA Format)

ACETRM-MPF
03/09



NOTE:

1. This booklet must be used in conjunction with the the data booklet for K-ACETRM, downloadable from where the product appears on the Megazyme website (www.megazyme.com).
2. Prepare the reagents and test samples as described in the data booklet for K-ACETRM.
2. For each batch of samples that are applied to the microplate format of K-ACETRM it is highly recommended that a standard calibration curve is included on the same microplate.

EQUIPMENT (RECOMMENDED):

1. Disposable 96 well polystyrene clear, flat bottom microplates e.g. Matrix Technologies Corp. cat. no. 4915 (www.matrixtechcorp.com).
2. Disposable 25 mL reagent reservoirs, e.g. Matrix Technologies Corp. cat. no. 8093 (www.matrixtechcorp.com).
3. Multichannel Micro-pipettors, e.g. Gilson Pipetman[®] Ultra 8-channel (1-20 μ L and 20-300 μ L).
4. Stop clock.
5. Microplate shaker, e.g. Heidolph Titramax 100 or 1000 (www.heidolph-instruments.com).
6. Microplate reader set at 340 nm.

MICROPLATE FORMAT:

- Wavelength:** 340 nm
Microplate: 96-well (e.g. clear flat-bottomed, glass or plastic)
Temperature: ~ 25°C
Final Volume: 0.266 µL
Sample Solution: 0.03-2.5 µg acetic acid per well
(in 10 µL sample volume)

Pipette into wells	Blank	Sample
distilled water (~25°C)	210 µL	200 µL
sample	-	10 µL
solution 2 (NADH/ATP/PEP/PVP buffer)	50 µL	50 µL
solution 3 (CoA)	2 µL*	2 µL*
suspension 4 (D-LDH/PTA/PK)	2 µL*	2 µL*
Mix**, read the absorbances of the solutions (A ₁) after approx. 2 min and start the reaction by addition of:		
Suspension 5 (AK)	2 µL*	2 µL*
Mix**, read the absorbances of the solutions (A ₂) at the end of the reaction (approx. 4 min). If the reaction has not stopped after 4 min, continue to read the absorbances at 2 min intervals until the absorbances increase constantly over 2 min.		

* if preferred, dilute sufficient enzyme for the required set of assays 1 in 5 with distilled water, and add 10 µL. Reduce the amount of water appropriately to maintain the same final volume.

** for example using microplate shaker, shake function on a microplate reader, or repeated aspiration (e.g. using pipettor set at 50 - 100 µL volume).

CALCULATION:

Calculations can be performed as described in the K-ACETRM data booklet* after appropriate path-length adjustment to 10 mm. This can either be performed automatically by the plate reader, or after manual determination of the true path-length (i.e. by simply performing a “manual” format assay of the standard solution in a 10 mm cuvette, and comparing the absorbance change to that of a reaction performed according to the “microplate” format). Alternatively a standard calibration curve can be used.

NOTE: Where sample readings can be corrected to a 10 mm path-length the calculations can be simplified by using the Megazyme *Mega-Calc*[™] *.

* available where the product appears on the Megazyme website (www.megazyme.com).



Megazyme International Ireland Ltd.,

Bray Business Park, Bray,

Co. Wicklow,

IRELAND

Telephone : (353.1) 286 1220

Fax : (353.1) 286 1264

Internet : www.megazyme.com

E-Mail : info@megazyme.com

WITHOUT GAURANTEE

The information contained in this booklet is, to the best of our knowledge, true and accurate, but since the conditions of use are beyond our control, no warranty is given or is implied in respect of any recommendation or suggestions which may be made or that any use will not infringe any patents.