

**Megazyme,**

Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.

Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Hydrogen Peroxide Assay Kit (Megaplex Red)
Cat. No: K-MRH2O2
Lot. No: 210437
Expiry Date: April, 2023

Bottle 1 210301
Bottle 2 210401
Bottle 3 210401B
Bottle 4 210201A
Bottle 5 210301A

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 5%	Corresponds
Blank reaction absorbance	+/- 10% of mean blank value	Corresponds

This lot meets our specification.

Signature: *Sarah Hurley*

Date: 09-Mar-2022

Sarah Hurley
Senior Biochemical QC Analyst

p.p. Ida Lazewska QC/QA Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)