

**Megazyme,**

Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.

Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Raffinose/D-Galactose Assay Kit
Cat. No: K-RAFGA
Lot. No: 210834
Expiry Date: December, 2023

Bottle 1 201101
Bottle 2 200201
Bottle 3 160201C
Bottle 4 170303A
Bottle 5 210528
Bottle 6 160906C

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2%	Corresponds
Blank reaction absorbance	+/- 10% of mean blank value	Corresponds

This lot meets our specification.

Signature:

Date: 31/08/2021

Elaine Delaney
Quality Assurance Specialist

p.p. Ida Lazewska QC/QA Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)