

**Megazyme,**

Bray Business Park,  
Bray, Co. Wicklow,  
A98 YV29,  
Ireland.

Tel: + 353 1 2861220 Fax: + 353 1 2861264

## Certificate of Analysis

**Product:** D-Xylose Assay Kit  
**Cat. No:** K-XYLOSE  
**Lot. No:** 220120  
**Expiry Date:** December, 2024

Bottle 1 211101  
Bottle 2 190502  
Bottle 3 190103  
Bottle 4 200401  
Bottle 5 141005c

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2%	Corresponds
Blank reaction absorbance	+/- 10% of mean blank value	Corresponds

This lot meets our specification.

**Signature:** *Sarah Hurley*

**Date:** 18-Jan-2022

Sarah Hurley  
Senior Biochemical QC Analyst

p.p. Ida Lazewska QC/QA Laboratory Manager

*(as this certificate is a computer printout, it has not been signed or dated by hand)*

*as not been signed or dated by hand)*