



Megazyme

Megazyme Building 1,
Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Inulotriose
Cat. No: O-INU3
Lot. No: 161003A
Expiry Date: N/A

Parameters:

| Purity By | Specification | Result |
|-----------|---------------|--------|
| HPLC | >85% | 88.1% |

This lot meets our specification.

Signature: 

Date: 18.JAN.2023

Ciara McLoughlin

QC Specialist

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme

Megazyme Building 1,
Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

HPLC:

Product: O-INU3, Lot 161003A

Column: Shodex Asahipak NH2P-40 3E analytical column

Temperature: 35°C

Flow rate: 0.35 mL/min (Eluent gradient shown below)

Detector: CAD (Charged Aerosol Detector)

HPLC System: Thermofisher U3000 Ultimate and Chromeleon v 7.0 software

| Time (min) | H ₂ O (%) | CH ₃ CN (%) |
|------------|----------------------|------------------------|
| 0 | 40 | 60 |
| 1 | 40 | 60 |
| 14 | 60 | 40 |
| 16 | 40 | 60 |
| 18 | 40 | 60 |
| 20 | 40 | 60 |

